


# Common Reporting Tool 1.4 (CRT)

GENERATE YOUR KAYE VALIDATOR<sup>®</sup> AVS  
AND VALPROBE<sup>®</sup> RT REPORTS WITH ONE  
DEDICATED TOOL



The Common Reporting Tool software is a comprehensive reporting utility that can be installed on a separate Windows PC or the Kaye Validation Console. It allows generating reports from Kaye Validator AVS or ValProbe RT Qualification study files to document the results of your validation studies. The Common Reporting Tool is seamlessly embedded in the validation console software suite. Qualification study files created on your Kaye Validator AVS or ValProbe RT console and stored on the console and/or shared network drive can also be transferred to the computer with the Common Reporting Tool via the sync out feature for report generation.

The ability to generate reports off-tablet on a separate PC adds flexibility and efficiency in the use of the Kaye consoles. All files are encrypted, tamper-proof, and can only be accessed by the CRT software, ensuring you maintain the same data integrity as our Validation Console, but with the added comfort of working in your own office.

The interface mirrors the patented Asset Centric Concept featured on Kaye Validation Consoles to ensure a seamless transition for users. Included in the CRT software is support for Microsoft's Active Directory\*.

Active Directory streamline user management by ensuring users are assigned the appropriate permissions and access. A free 30-day trial is available for download on the Kaye website. A base license can then be purchased for any number of PCs at your site that would like to view and manage the reporting data.

### BENEFITS

- View data and generate reports from a Validator AVS or ValProbe RT system at the convenience of your desk or home office
- Read, manage, and store data on your secured network or server, allowing for easy access to files


- Merging of up to three Kaye Validator AVS or ValProbe RT studies into one single report
- Detailed Pass/Fail Report, including info on passed qualifications
- Report for climatic and thermostatic chambers in compliance with AFNOR FD X15-140 or IEC 60068 normative documents for ValProbe RT
- Encrypted and tamper-proof files can only be read through Common Reporting Tool software, ensuring data integrity
- Incorporating Microsoft's Active Directory, ensuring seamless and convenient user management

**FD X15-140 Report**  
Study Name: Min. load product  
Printed on 21-Mar-2024 at 11:33:30 by JDS

Temperature Data (°C)

Sensor Label	SN	Min	Max	Max - Min	Std Dev	Uncertainty	Avg	Unit	Specification	Avg ± UMI	Calc. Spec. Max	Avg - UMI	Calc. Spec. Min	Result
Sensor17	TE19-A	20.05	21.15	0.23	0.06	0.40	21.05	0.42	22.00	21.47	25.00	20.63	20.00	PASS
Sensor17	1151-A	21.42	21.94	0.52	0.17	0.30	21.70	0.45	22.00	22.15	25.00	21.25	20.00	PASS
Sensor13	UG97-A	20.88	21.22	0.34	0.08	0.20	20.94	0.26	22.00	21.20	25.00	20.68	20.00	PASS
Sensor14	UZ99-B	21.48	21.82	0.34	0.10	0.10	21.63	0.22	22.00	21.85	25.00	21.41	20.00	PASS
Sensor15	M275-B	21.28	21.53	0.25	0.07	0.40	21.30	0.42	22.00	21.81	25.00	20.97	20.00	PASS
Sensor7	SH15-A	20.97	21.05	0.08	0.02	0.40	21.00	0.40	22.00	21.40	25.00	20.60	20.00	PASS
Sensor3	UG98-A	20.93	20.98	0.05	0.01	0.40	20.95	0.40	22.00	21.35	25.00	20.55	20.00	PASS
Sensor4	UG98-A	21.14	21.24	0.10	0.03	0.40	21.18	0.40	22.00	21.58	25.00	20.78	20.00	PASS
Sensor15	UZ99-A	21.16	21.26	0.10	0.03	0.20	21.20	0.21	22.00	21.41	25.00	20.99	20.00	PASS
Sensor19	MA15-A	21.21	21.30	0.15	0.04	0.20	21.28	0.22	22.00	21.50	25.00	21.06	20.00	PASS
Sensor5	UH95-A	20.83	21.11	0.28	0.06	0.20	20.90	0.23	22.00	21.13	25.00	20.67	20.00	PASS
Sensor1	RP92-A	21.16	21.28	0.12	0.04	0.30	21.22	0.31	22.00	21.53	25.00	20.91	20.00	PASS
Sensor18	PL41-T	20.87	21.38	0.51	0.13	0.30	21.07	0.39	22.00	21.46	25.00	20.68	20.00	PASS
Sensor20	UZ99-A	21.26	21.48	0.20	0.06	0.30	21.35	0.32	22.00	21.67	25.00	21.03	20.00	PASS
Sensor7	M275-A	21.53	21.45	0.12	0.03	0.30	21.58	0.31	22.00	21.69	25.00	21.07	20.00	PASS
Sensor11	MA15-R	21.35	21.46	0.11	0.03	0.20	21.40	0.21	22.00	21.61	25.00	21.19	20.00	PASS
Sensor19	TW93-A	21.29	21.39	0.10	0.03	0.20	21.33	0.21	22.00	21.54	25.00	21.12	20.00	PASS
Sensor11	UH95-A	21.01	21.05	0.04	0.01	0.20	21.03	0.20	22.00	21.23	25.00	20.83	20.00	PASS
Sensor10	UH98-R	21.18	21.32	0.14	0.04	0.20	21.24	0.21	22.00	21.45	25.00	21.03	20.00	PASS

Climatic chamber data report



Climatic chamber Graph report

### Kaye Representative contact:

**EUROPE, MIDDLE EAST, AFRICA AND ASIA**  
Amphenol Advanced Sensors Germany GmbH  
Sinzheimer Strasse 6  
D-75179 Pforzheim  
T: +49 (0) 7231-14 335 0  
F: +49 (0) 7231-14335 29  
Email: kaye@amphenol-sensors.com

**USA/AMERICAS**  
Amphenol Thermometrics, Inc.  
967 Windfall Road  
St. Marys, PA 15857  
T: +1(814) 834-9140  
F: +1(814) 781-7969  
Email: kaye-us@amphenol-sensors.com

### Request a demo:

**INDIA**  
Amphenol Interconnect India Pvt Ltd.  
Plot no. 6, Survey No.64 | Software Units layout  
MAHAVEER TECHNO PARK  
Hitech City, Madhapur | Hyderabad,  
Telangana – 500081 | T: +91 40 33147100  
Email: kaye-india@amphenol-sensors.com

**CHINA**  
Amphenol (Changzhou) Connector Systems Co., Ltd, Building 10,  
Jintong Industrial Park, No. 8 Xihu Road,  
Wujin High-Tech Development Zone,  
Changzhou, Jiangsu 213164  
T: 0086-519-83055197



**Warranty and disclaimer:** The information mentioned on documents are based on our current tests, knowledge and experience. Because of the effect of possible influences in an application of the product, they do not exempt the user from their own tests, checks and trials. A guarantee of certain properties or a guarantee for the proper suitability of the product for a specific, especially permanent application can not be derived from our data. Liability is therefore excluded to that extent permitted by law. Any proprietary rights of third parties as well as existing laws and regulations must be observed by the recipient of the product on his own responsibility.

© 2024 Amphenol Corporation. All Rights Reserved. Specifications are subject to change without notice. Other company names and product names used in this document are the registered trademarks or trademarks of their respective owners.